

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112919\
 Data File : VX013713.D
 Acq On : 29 Nov 2019 19:53
 Operator : JC/SP
 Sample : K5965-01DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S39-0187DL

Quant Time: Dec 02 07:28:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	674698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1030831	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	910382	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	421634	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	368623	47.07	ug/l	0.00
Spiked Amount			Recovery	=	94.14%	
35) Dibromofluoromethane	5.49	113	304085	47.72	ug/l	0.00
Spiked Amount			Recovery	=	95.44%	
50) Toluene-d8	8.71	98	1220280	49.15	ug/l	0.00
Spiked Amount			Recovery	=	98.30%	
62) 4-Bromofluorobenzene	11.14	95	411659	45.98	ug/l	0.00
Spiked Amount			Recovery	=	91.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	5667	0.598	ug/l #	86
27) cis-1,2-Dichloroethene	4.59	96	442554	55.093	ug/l	89
31) Cyclohexane	5.57	56	44730	3.866	ug/l	95
39) Methylcyclohexane	7.46	83	98901	8.551	ug/l	98
40) Benzene	6.13	78	42857	1.504	ug/l	98
67) Ethyl Benzene	10.25	91	205887	6.339	ug/l	100
68) m/p-Xylenes	10.36	106	158681	12.761	ug/l	100
69) o-Xylene	10.70	106	20225	1.668	ug/l	94
73) Isopropylbenzene	11.01	105	42526	1.438	ug/l	100
78) n-propylbenzene	11.35	91	32862	0.995	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	240439	9.773	ug/l	99
85) sec-Butylbenzene	11.94	105	10196	0.362	ug/l #	67
86) p-Isopropyltoluene	12.06	119	13701	0.523	ug/l	86
95) Naphthalene	13.83	128	87437	3.061	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112919\
Data File : VX013713.D
Acq On : 29 Nov 2019 19:53
Operator : JC/SP
Sample : K5965-01DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S39-0187DL

Quant Time: Dec 02 07:28:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

