

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
 Data File : VX044081.D
 Acq On : 03 Dec 2024 10:11
 Operator : JC/MD
 Sample : VX1203MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1203MBL01

Quant Time: Dec 03 23:38:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	117409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	223839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	197050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85483	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	102184	50.743	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.480%
35) Dibromofluoromethane	5.379	113	75771	47.373	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.740%
50) Toluene-d8	8.646	98	273530	49.612	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.220%
62) 4-Bromofluorobenzene	11.079	95	93391	49.772	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
Data File : VX044081.D
Acq On : 03 Dec 2024 10:11
Operator : JC/MD
Sample : VX1203MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1203MBL01

Quant Time: Dec 03 23:38:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

