

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
 Data File : VX044082.D
 Acq On : 03 Dec 2024 10:34
 Operator : JC/MD
 Sample : VX1203WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1203WBL01

Quant Time: Dec 03 23:38:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	113150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	222913	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	189304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	102003	52.559	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.120%	
35) Dibromofluoromethane	5.379	113	74150	46.552	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	93.100%	
50) Toluene-d8	8.647	98	271717	49.487	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.980%	
62) 4-Bromofluorobenzene	11.079	95	92538	49.522	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	99.040%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
Data File : VX044082.D
Acq On : 03 Dec 2024 10:34
Operator : JC/MD
Sample : VX1203WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1203WBL01

Quant Time: Dec 03 23:38:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

