

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
 Data File : VX044086.D
 Acq On : 03 Dec 2024 12:10
 Operator : JC/MD
 Sample : P4960-05
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW3

Quant Time: Dec 03 23:41:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	118492	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	228851	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	93386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104344	51.342	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.680%			
35) Dibromofluoromethane	5.379	113	72192	44.147	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 88.300%			
50) Toluene-d8	8.647	98	283276	50.254	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.500%			
62) 4-Bromofluorobenzene	11.079	95	101678	53.002	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 106.000%			
Target Compounds						
95) Naphthalene	13.780	128	1962	0.294	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
Data File : VX044086.D
Acq On : 03 Dec 2024 12:10
Operator : JC/MD
Sample : P4960-05
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW3

Quant Time: Dec 03 23:41:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

