

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
 Data File : VX044095.D
 Acq On : 03 Dec 2024 16:04
 Operator : JC/MD
 Sample : P5006-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C6-2

Quant Time: Dec 04 00:10:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	112488	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	223144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	201427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87397	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	102962	53.366	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.740%		
35) Dibromofluoromethane	5.379	113	60667	38.048	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 76.100%		
50) Toluene-d8	8.647	98	278115	50.600	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.200%		
62) 4-Bromofluorobenzene	11.079	95	97639	52.198	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 104.400%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	15777	17.804	ug/l	95
67) Ethyl Benzene	10.195	91	26401	3.522	ug/l	97
69) o-Xylene	10.647	106	2985	1.094	ug/l	93
84) 1,2,4-Trimethylbenzene	11.750	105	13956	2.536	ug/l	98
95) Naphthalene	13.774	128	1173211	187.672	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
Data File : VX044095.D
Acq On : 03 Dec 2024 16:04
Operator : JC/MD
Sample : P5006-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C6-2

Quant Time: Dec 04 00:10:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

