

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
 Data File : VX044096.D
 Acq On : 03 Dec 2024 16:27
 Operator : JC/MD
 Sample : P5006-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C6-3

Quant Time: Dec 04 00:10:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	118097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	224861	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	199085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	90167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	106228	52.444	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.880%			
35) Dibromofluoromethane	5.379	113	59586	37.085	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 74.160%#			
50) Toluene-d8	8.647	98	273575	49.394	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.780%			
62) 4-Bromofluorobenzene	11.079	95	99076	52.562	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.120%			
Target Compounds						
16) Acetone	2.380	43	16316	17.538	ug/l	99
27) cis-1,2-Dichloroethene	4.507	96	1762	1.015	ug/l #	6
67) Ethyl Benzene	10.195	91	62654	8.456	ug/l	97
69) o-Xylene	10.640	106	5145	1.907	ug/l	92
73) Isopropylbenzene	10.964	105	8444	1.218	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	6996	1.227	ug/l	93
84) 1,2,4-Trimethylbenzene	11.750	105	28947	5.099	ug/l	99
95) Naphthalene	13.774	128	2120679	328.812	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
Data File : VX044096.D
Acq On : 03 Dec 2024 16:27
Operator : JC/MD
Sample : P5006-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C6-3

Quant Time: Dec 04 00:10:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

