

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
 Data File : VX044103.D
 Acq On : 03 Dec 2024 19:09
 Operator : JC/MD
 Sample : P5006-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C12-1

Quant Time: Dec 04 00:15:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	109453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	218071	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	193451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	89950	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	100954	53.776	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.560%			
35) Dibromofluoromethane	5.385	113	52285	33.554	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 67.100%#			
50) Toluene-d8	8.647	98	270384	50.338	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.680%			
62) 4-Bromofluorobenzene	11.079	95	96584	52.835	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.680%			
Target Compounds						
16) Acetone	2.386	43	27317	31.682	ug/l	98
27) cis-1,2-Dichloroethene	4.477	96	4252	2.642	ug/l	80
40) Benzene	6.031	78	13245	2.190	ug/l	94
52) Toluene	8.720	92	8980	2.481	ug/l	97
67) Ethyl Benzene	10.189	91	154342	21.438	ug/l	99
68) m/p-Xylenes	10.299	106	17341	6.459	ug/l	100
69) o-Xylene	10.640	106	20660	7.881	ug/l	99
73) Isopropylbenzene	10.963	105	14094	2.038	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	14658	2.577	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	53435	9.435	ug/l	98
95) Naphthalene	13.774	128	3375861	524.690	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
Data File : VX044103.D
Acq On : 03 Dec 2024 19:09
Operator : JC/MD
Sample : P5006-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C12-1

Quant Time: Dec 04 00:15:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

