

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
 Data File : VX044107.D
 Acq On : 03 Dec 2024 20:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 04 00:17:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.633	0.551	13.0	64	0.00
3 P	Chloromethane	0.667	0.586	12.1	67	0.00
4 C	Vinyl Chloride	0.709	0.661	6.8#	69	0.00
5 T	Bromomethane	0.436	0.463	-6.2	82	0.00
6 T	Chloroethane	0.432	0.426	1.4	74	0.00
7 T	Trichlorofluoromethane	1.268	1.250	1.4	73	0.00
8 T	Diethyl Ether	0.462	0.491	-6.3	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.605	0.588	2.8	71	0.00
10 T	Methyl Iodide	0.755	0.773	-2.4	76	0.00
11 T	Tert butyl alcohol	0.138	0.143	-3.6	78	0.00
12 CM	1,1-Dichloroethene	0.576	0.562	2.4#	74	0.00
13 T	Acrolein	0.145	0.177	-22.1	97	0.00
14 T	Allyl chloride	0.941	1.030	-9.5	78	0.00
15 T	Acrylonitrile	0.335	0.366	-9.3	78	0.00
16 T	Acetone	0.394	0.357	9.4	67	0.00
17 T	Carbon Disulfide	1.188	0.933	21.5	57	0.00
18 T	Methyl Acetate	0.917	1.087	-18.5	87	0.00
19 T	Methyl tert-butyl Ether	2.092	2.418	-15.6	85	0.00
20 T	Methylene Chloride	0.668	0.669	-0.1	80	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.585	1.7	73	0.00
22 T	Diisopropyl ether	2.013	2.311	-14.8	85	0.00
23 T	Vinyl Acetate	1.722	1.940	-12.7	79	0.00
24 P	1,1-Dichloroethane	1.161	1.262	-8.7	80	0.00
25 T	2-Butanone	0.508	0.542	-6.7	76	0.00
26 T	2,2-Dichloropropane	1.049	1.058	-0.9	73	0.00
27 T	cis-1,2-Dichloroethene	0.735	0.773	-5.2	77	0.00
28 T	Bromochloromethane	0.511	0.531	-3.9	93	0.01
29 T	Tetrahydrofuran	0.312	0.341	-9.3	80	0.00
30 C	Chloroform	1.249	1.396	-11.8#	84	0.00
31 T	Cyclohexane	0.956	0.874	8.6	67	0.00
32 T	1,1,1-Trichloroethane	1.077	1.191	-10.6	80	0.00
33 S	1,2-Dichloroethane-d4	0.858	0.889	-3.6	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.357	0.326	8.7	89	0.00
36 T	1,1-Dichloropropene	0.445	0.415	6.7	75	0.00
37 T	Ethyl Acetate	0.553	0.565	-2.2	82	0.00
38 T	Carbon Tetrachloride	0.500	0.470	6.0	75	0.00
39 T	Methylcyclohexane	0.578	0.487	15.7	66	0.00
40 TM	Benzene	1.387	1.307	5.8	78	0.00
41 T	Methacrylonitrile	0.283	0.306	-8.1	83	0.00
42 TM	1,2-Dichloroethane	0.557	0.578	-3.8	85	0.00
43 T	Isopropyl Acetate	0.880	0.927	-5.3	84	0.00
44 TM	Trichloroethene	0.393	0.326	17.0	78	0.00
45 C	1,2-Dichloropropane	0.340	0.347	-2.1#	83	0.00
46 T	Dibromomethane	0.269	0.279	-3.7	85	0.00
47 T	Bromodichloromethane	0.482	0.495	-2.7	81	0.00
48 T	Methyl methacrylate	0.422	0.436	-3.3	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
 Data File : VX044107.D
 Acq On : 03 Dec 2024 20:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 04 00:17:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	84	0.00
50 S	Toluene-d8	1.232	1.101	10.6	89	0.00
51 T	4-Methyl-2-Pentanone	0.537	0.584	-8.8	86	0.00
52 CM	Toluene	0.830	0.807	2.8#	78	0.00
53 T	t-1,3-Dichloropropene	0.491	0.519	-5.7	81	0.00
54 T	cis-1,3-Dichloropropene	0.539	0.544	-0.9	79	0.00
55 T	1,1,2-Trichloroethane	0.343	0.350	-2.0	84	0.00
56 T	Ethyl methacrylate	0.534	0.584	-9.4	85	0.00
57 T	1,3-Dichloropropane	0.577	0.597	-3.5	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.273	0.292	-7.0	85	0.00
59 T	2-Hexanone	0.403	0.436	-8.2	82	0.00
60 T	Dibromochloromethane	0.344	0.366	-6.4	82	0.00
61 T	1,2-Dibromoethane	0.346	0.363	-4.9	84	0.00
62 S	4-Bromofluorobenzene	0.419	0.409	2.4	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.330	0.309	6.4	77	0.00
65 PM	Chlorobenzene	1.098	1.039	5.4	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.356	0.378	-6.2	83	0.00
67 C	Ethyl Benzene	1.861	1.865	-0.2#	80	0.00
68 T	m/p-Xylenes	0.694	0.675	2.7	77	0.00
69 T	o-Xylene	0.678	0.673	0.7	77	0.00
70 T	Styrene	1.121	1.160	-3.5	80	0.00
71 P	Bromoform	0.240	0.247	-2.9	76	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.843	3.717	3.3	80	0.00
74 T	N-amyl acetate	1.780	1.925	-8.1	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.316	1.333	-1.3	84	0.00
76 T	1,2,3-Trichloropropane	1.096	1.099	-0.3	82	0.00
77 T	Bromobenzene	0.917	0.881	3.9	79	0.00
78 T	n-propylbenzene	4.293	4.187	2.5	78	0.00
79 T	2-Chlorotoluene	2.714	2.600	4.2	80	0.00
80 T	1,3,5-Trimethylbenzene	3.162	3.120	1.3	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.402	0.369	8.2	73	0.00
82 T	4-Chlorotoluene	3.064	3.015	1.6	79	0.00
83 T	tert-Butylbenzene	3.134	3.088	1.5	79	0.00
84 T	1,2,4-Trimethylbenzene	3.148	3.178	-1.0	81	0.00
85 T	sec-Butylbenzene	3.840	3.694	3.8	76	0.00
86 T	p-Isopropyltoluene	3.119	3.142	-0.7	78	0.00
87 T	1,3-Dichlorobenzene	1.670	1.601	4.1	80	0.00
88 T	1,4-Dichlorobenzene	1.702	1.627	4.4	80	0.00
89 T	n-Butylbenzene	2.750	2.744	0.2	76	0.00
90 T	Hexachloroethane	0.530	0.490	7.5	73	0.00
91 T	1,2-Dichlorobenzene	1.685	1.615	4.2	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.264	0.274	-3.8	85	0.00
93 T	1,2,4-Trichlorobenzene	0.994	0.942	5.2	76	0.00
94 T	Hexachlorobutadiene	0.393	0.352	10.4	75	0.00
95 T	Naphthalene	3.576	4.234	-18.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120324\
Data File : VX044107.D
Acq On : 03 Dec 2024 20:41
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 04 00:17:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.006	0.997	0.9	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6